

Application Memo

Industry Inorganic chemical industry
Instrument
Measurement method
Standards

After adding 10% potassium iodide and acetic acid to the sample including Cu^{2+} , it is titrated with 0.1 mol / L sodium thiosulfate solution. The endpoint is the maximum inflexion on the titration curve. The copper ion concentration is calculated from the titration volume.

Main unit Automatic potentiometric titrator (preamplifier STD)
Electrode Platinum electrode
 C